

## Liberty University AED Pre-Shift Inspection Checklist

Employee Name: \_\_\_\_\_ Employee ID: \_\_\_\_\_

AED Serial Number: \_\_\_\_\_ Date/Time: \_\_\_\_\_

Inspection	YES	NO	N/A
Is the light flashing green?			
Is there an in-date Adult PAD loaded into the AED?			
Is there an in-date Child/Infant PAD with the AED?			
Does the cabinet impede egress or cause an ADA issue?			
Is there a complete, Fast Response Kit with the AED?			
Does the door alarm sound upon opening?			
Is the alarm pin in the cabinet set to the correct position to prevent false alarms?			
Is the cabinet in good shape (not scratches, broken glass, broken door, etc.)?			